



Kikkoman Biochemifa Company

**CERTIFICATE OF ANALYSIS**

**LuciPac W**

Product code :61303

Lot No.: 20180905L

Expiry date: Dec. 04, 2019

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	11600	$\geq 9100, \leq 16900$
(When applied 100 $\mu$ L of 5pmol AMP: measured with Lumitester PD-10N)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hiroshi Katayama, Manager

Date Sep. 12, 2018

KBG6032901aa-Eaa  
SA1809-0085

Quality Control Group  
Edogawa Plant  
Kikkoman Biochemifa Company