



CERTIFICATE OF ANALYSIS

LuciPac W

Product code :61303

Lot No.: 20180906L

Expiry date: Dec. 05, 2019

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	11000	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5pmol AMP: measured with Lumitester PD-10N)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H Katayama
Hiroshi Katayama; Manager

Date Oct. 19, 2018

KBG6032901aa-Eaa
SA1810-0110

Quality Control Group
Edogawa Plant
Kikkoman Biochemifa Company