



**CERTIFICATE OF ANALYSIS**

**LuciPac W**

Product code :61303

Lot No.: 20181220L

Expiry date: Mar. 19, 2020

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	11200	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5pmol AMP: measured with Lumitester PD-10N)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama  
Hiroshi Katayama; Manager

Date Jan. 17, 2019

KBG6032901aa-Eaa  
SA1901-0039

Quality Control Group  
Edogawa Plant  
Kikkoman Biochemifa Company