



CERTIFICATE OF ANALYSIS

LuciPac W

Product code :61303

Lot No.: 20190301L

Expiry date: May. 31, 2020

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|---------------------------|---------------|-------------------------|
| Average Luminescence(RLU) | 10500 | $\geq 9100, \leq 16900$ |

(When applied 100 μ L of 5pmol AMP: measured with Lumitester PD-10N)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H Katayama
Hiroshi Katayama; Manager

Date Mar. 11, 2019

KBG6032901aa-Eaa
SA1903-0084

Quality Control Group
Edogawa Plant Production Department
Kikkoman Biochemifa Company