



**CERTIFICATE OF ANALYSIS**

**LuciPac W**

Product code :61303

---

Lot No.: 20190422L

Expiry date: Jul. 21, 2020

---

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
-----------------------	---------------	----------------------

Average Luminescence(RLU)	11300	$\geq 9100, \leq 16900$
---------------------------	-------	-------------------------

(When applied 100 $\mu$ L of 5pmol AMP: measured with Lumitester PD-10N)

---

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H Katayama  
Hiroshi Katayama, Manager

Date May. 14, 2019

KBG6032901aa-Eaa  
SA1905-0068

Quality Control Group  
Edogawa Plant Production Department  
Kikkoman Biochemifa Company