



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

October 3, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Wheat/Gluten (Gliadin) ELISA Kit II	19OCSFGD070	OCT. 02 2020

The product kit, batch 19OCSFGD070, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 3, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.005	0.005	0.005	0.005	9.5
0.78	0.036	0.039	0.038	0.036	0.037	4.0
1.56	0.072	0.071	0.072	0.069	0.071	2.0
3.12	0.141	0.136	0.136	0.136	0.137	1.8
6.25	0.284	0.275	0.276	0.272	0.277	1.9
12.5	0.568	0.556	0.548	0.547	0.555	1.8
25	1.115	1.098	1.089	1.074	1.094	1.6
50	1.990	1.976	1.936	1.897	1.950	2.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.5	6.6-12.2
B	8.1	5.6-10.4
C	9.5	7.0-13.1