

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Wheat/Gluten (Gliadin) ELISA Kit II	19JUSFGD067	JUN. 05 2020

The product kit, batch 19JUSFGD067, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 06, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.010	0.012	0.010	0.011	10.5
0.78	0.047	0.042	0.043	0.042	0.044	5.5
1.56	0.078	0.074	0.074	0.073	0.075	3.0
3.12	0.142	0.133	0.134	0.133	0.136	3.2
6.25	0.275	0.261	0.258	0.260	0.264	2.9
12.5	0.512	0.511	0.510	0.504	0.509	0.7
25	1.041	1.006	0.995	0.987	1.007	2.4
50	1.857	1.811	1.801	1.790	1.815	1.6

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.9	6.6-12.2
B	8.3	5.6-10.4
C	9.4	7.0-13.1