



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

December 23, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
High Sensitive Peanut ELISA Kit II	21DESFHP012	DEC.22 2022

The product kit, batch 21DESFHP012, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : December 23, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.037	0.038	0.036	0.041	0.038	
0.5	0.078	0.081	0.080	0.079	0.080	1.6
1	0.118	0.116	0.121	0.118	0.118	1.7
2	0.199	0.200	0.202	0.207	0.202	1.8
4	0.347	0.356	0.346	0.358	0.352	1.7
8	0.631	0.635	0.636	0.647	0.637	1.1
16	1.210	1.207	1.209	1.217	1.211	0.4
32	2.241	2.257	2.246	2.250	2.249	0.3

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.5	5.4-9.9
B	6.9	6.1-11.4
C	6.3	5.5-10.3