



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

October 13, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Peanut ELISA Kit II	22OCSFPE068	OCT.11 2023

The product kit, batch 22OCSFPE068, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 13, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.010	0.011	0.009	0.010	
0.78	0.052	0.052	0.054	0.049	0.052	4.0
1.56	0.097	0.094	0.094	0.088	0.093	4.0
3.13	0.174	0.167	0.167	0.163	0.168	2.7
6.25	0.324	0.313	0.310	0.302	0.312	2.9
12.5	0.608	0.591	0.593	0.589	0.595	1.5
25	1.147	1.126	1.113	1.134	1.130	1.3
50	2.030	1.991	2.007	2.030	2.015	0.9

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.7	4.7-8.7
B	6.7	5.8-10.8
C	3.9	3.0-5.6