



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

July 20, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Peanut ELISA Kit II	21JLSFPE059	JUL.19 2022

The product kit, batch 21JLSFPE059, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : July 20, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.013	0.015	0.011	0.010	0.012	
0.78	0.054	0.052	0.053	0.053	0.053	1.5
1.56	0.095	0.090	0.091	0.090	0.092	2.6
3.13	0.178	0.172	0.177	0.175	0.176	1.5
6.25	0.335	0.328	0.325	0.329	0.329	1.3
12.5	0.633	0.615	0.622	0.630	0.625	1.3
25	1.177	1.136	1.183	1.176	1.168	1.8
50	2.098	2.072	2.108	2.073	2.088	0.9

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	5.4	4.5-8.4
B	6.4	5.7-10.6
C	4.0	2.7-5.1