



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 12, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Peanut ELISA Kit II	21MYSFPE058	MAY.11 2022

The product kit, batch 21MYSFPE058, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 12, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.009	0.008	0.009	0.008	0.009	
0.78	0.047	0.046	0.044	0.044	0.045	3.3
1.56	0.085	0.084	0.083	0.082	0.084	1.5
3.13	0.163	0.162	0.160	0.155	0.160	2.2
6.25	0.307	0.303	0.299	0.290	0.300	2.4
12.5	0.582	0.574	0.574	0.561	0.573	1.5
25	1.099	1.096	1.072	1.064	1.083	1.6
50	1.944	1.982	1.945	1.910	1.945	1.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.8	5.0-9.3
B	8.6	6.4-11.8
C	4.7	2.6-4.9