



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

September 4, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Peanut ELISA Kit II	20SESFPE053	SEP.03 2021

The product kit, batch 20SESFPE053, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 4, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.010	0.008	0.008	0.009	
0.78	0.047	0.047	0.047	0.047	0.047	0.0
1.56	0.090	0.087	0.085	0.085	0.087	2.7
3.13	0.170	0.167	0.164	0.162	0.166	2.1
6.25	0.318	0.311	0.310	0.310	0.312	1.2
12.5	0.606	0.594	0.592	0.598	0.598	1.0
25	1.152	1.144	1.115	1.111	1.131	1.8
50	2.072	2.010	2.029	2.025	2.034	1.3

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.5	5.0-9.3
B	6.6	6.4-11.8
C	3.3	2.6-4.9