



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

April 20, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Peanut ELISA Kit II	20APSFPE051	APR.18 2021

The product kit, batch 20APSFPE051, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : April 20, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.009	0.008	0.009	0.009	
0.78	0.051	0.051	0.048	0.049	0.050	3.0
1.56	0.096	0.093	0.088	0.089	0.092	4.0
3.12	0.186	0.179	0.177	0.174	0.179	2.8
6.25	0.353	0.343	0.339	0.330	0.341	2.8
12.5	0.678	0.655	0.648	0.642	0.656	2.4
25	1.282	1.245	1.232	1.205	1.241	2.6
50	2.266	2.292	2.206	2.220	2.246	1.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.0	5.0-9.3
B	7.0	6.4-11.8
C	3.5	2.6-4.9