



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

April 11, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	24APSFOA162	APR.09 2025

The product kit, batch 24APSFOA162, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : April 11, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.013	0.014	0.013	0.012	0.013	
0.78	0.047	0.047	0.047	0.048	0.047	1.1
1.56	0.082	0.084	0.082	0.082	0.083	1.2
3.13	0.149	0.145	0.148	0.146	0.147	1.2
6.25	0.284	0.278	0.272	0.277	0.278	1.8
12.5	0.548	0.538	0.536	0.540	0.541	1.0
25	1.064	1.045	1.039	1.040	1.047	1.1
50	2.006	1.988	1.986	1.987	1.992	0.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.7	5.7-10.5
B	6.0	3.8-7.0
C	2.9	2.5-4.6