



## Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,  
Yokohama-shi, 230-8504, Japan  
TEL:+81-45-586-2514  
E-mail: info\_miobs\_e@morinaga.co.jp

January 12, 2024

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

*M. Morita*

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	24JASFOA160	JAN.10 2025

The product kit, batch 24JASFOA160, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : January 12, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.009	0.010	0.008	0.009	
0.78	0.048	0.046	0.046	0.044	0.046	3.5
1.56	0.081	0.083	0.079	0.080	0.081	2.1
3.13	0.148	0.149	0.148	0.146	0.148	0.9
6.25	0.297	0.289	0.290	0.284	0.290	1.8
12.5	0.596	0.586	0.578	0.563	0.581	2.4
25	1.103	1.094	1.060	1.062	1.080	2.0
50	2.075	2.110	2.053	2.035	2.068	1.6

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.8	5.7-10.5
B	6.3	3.8-7.0
C	3.1	2.5-4.6