



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

November 14, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	23NOSFOA158	NOV.12 2024

The product kit, batch 23NOSFOA158, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : November 14, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.011	0.011	0.011	0.011	
0.78	0.051	0.050	0.050	0.051	0.051	1.1
1.56	0.094	0.090	0.089	0.088	0.090	2.9
3.13	0.172	0.168	0.165	0.166	0.168	1.8
6.25	0.336	0.325	0.318	0.314	0.323	3.0
12.5	0.649	0.617	0.616	0.607	0.622	3.0
25	1.242	1.201	1.204	1.183	1.208	2.1
50	2.308	2.247	2.230	2.202	2.247	2.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.2	5.7-10.5
B	5.9	3.8-7.0
C	3.1	2.5-4.6