



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

October 20, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	23OCSFOA156	OCT.18 2024

The product kit, batch 23OCSFOA156, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 20, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.010	0.010	0.010	0.011	
0.78	0.053	0.050	0.051	0.052	0.052	2.5
1.56	0.094	0.092	0.089	0.090	0.091	2.4
3.13	0.171	0.168	0.167	0.166	0.168	1.3
6.25	0.340	0.322	0.323	0.326	0.328	2.5
12.5	0.653	0.627	0.623	0.631	0.634	2.1
25	1.220	1.204	1.181	1.194	1.200	1.4
50	2.184	2.239	2.274	2.228	2.231	1.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.5	5.7-10.5
B	5.8	3.8-7.0
C	2.9	2.5-4.6