



## Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,  
Yokohama-shi, 230-8504, Japan  
TEL:+81-45-586-2514  
E-mail: info\_miobs\_e@morinaga.co.jp

May 11, 2023

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

*M. Morita*

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	23MYSFOA149	MAY.10 2024

The product kit, batch 23MYSFOA149, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : May 11, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.011	0.011	0.010	0.011	
0.78	0.044	0.042	0.042	0.039	0.042	3.4
1.56	0.080	0.079	0.078	0.076	0.078	1.1
3.13	0.149	0.146	0.144	0.144	0.146	2.1
6.25	0.287	0.280	0.275	0.277	0.280	1.6
12.5	0.566	0.545	0.536	0.548	0.549	1.3
25	1.069	1.053	1.038	1.050	1.053	1.1
50	2.026	2.006	1.943	1.964	1.985	2.1

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.5	5.7-10.5
B	4.0	3.8-7.0
C	4.3	2.5-4.6