



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

August 3, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22AUSFOA138	AUG.02 2023

The product kit, batch 22AUSFOA138, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : August 3, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.020	0.020	0.022	0.020	0.021	
0.78	0.058	0.058	0.058	0.058	0.058	0.0
1.56	0.094	0.092	0.092	0.092	0.093	1.1
3.13	0.164	0.157	0.154	0.162	0.159	2.9
6.25	0.301	0.293	0.290	0.303	0.297	2.1
12.5	0.580	0.553	0.549	0.563	0.561	2.5
25	1.048	1.033	1.005	1.076	1.041	2.8
50	1.894	1.874	1.802	1.895	1.866	2.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.4	5.7-10.5
B	3.9	3.8-7.0
C	3.1	2.5-4.6