



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

July 25, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22JLSFOA137	JUL.22 2023

The product kit, batch 22JLSFOA137, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : July 25, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.026	0.025	0.026	0.026	0.026	
0.78	0.061	0.059	0.059	0.062	0.060	2.5
1.56	0.097	0.096	0.095	0.098	0.097	1.3
3.13	0.166	0.164	0.162	0.161	0.163	1.4
6.25	0.305	0.295	0.295	0.293	0.297	1.8
12.5	0.590	0.574	0.564	0.575	0.576	1.9
25	1.081	1.060	1.044	1.063	1.062	1.4
50	1.990	1.981	1.982	1.988	1.985	0.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.9	5.7-10.5
B	5.5	3.8-7.0
C	2.9	2.5-4.6