



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 23, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22MYSFOA135	MAY.20 2023

The product kit, batch 22MYSFOA135, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 23, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.013	0.006	0.010	0.010	0.010	
0.78	0.043	0.044	0.045	0.045	0.044	2.2
1.56	0.078	0.076	0.074	0.076	0.076	2.1
3.13	0.150	0.154	0.142	0.144	0.148	3.7
6.25	0.289	0.281	0.280	0.276	0.282	1.9
12.5	0.563	0.550	0.540	0.548	0.550	1.7
25	1.070	1.052	1.038	1.046	1.052	1.3
50	1.952	1.941	1.928	1.920	1.935	0.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.9	6.1-11.4
B	4.7	4.4-8.2
C	3.2	2.8-5.3