



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 10, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22MYSFOA134	MAY.09 2023

The product kit, batch 22MYSFOA134, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 10, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.008	0.011	0.009	0.010	
0.78	0.051	0.051	0.049	0.051	0.051	2.0
1.56	0.088	0.087	0.085	0.084	0.086	2.1
3.13	0.162	0.162	0.160	0.162	0.162	0.6
6.25	0.306	0.306	0.296	0.303	0.303	1.6
12.5	0.588	0.595	0.591	0.579	0.588	1.2
25	1.073	1.098	1.068	1.075	1.079	1.2
50	1.973	2.027	2.009	2.004	2.003	1.1

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.9	6.1-11.4
B	5.2	4.4-8.2
C	3.5	2.8-5.3