



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

April 22, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22APSFOA133	APR.20 2023

The product kit, batch 22APSFOA133, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : April 22, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.011	0.011	0.011	0.011	
0.78	0.052	0.052	0.051	0.050	0.051	1.9
1.56	0.089	0.089	0.089	0.087	0.089	1.1
3.13	0.171	0.165	0.164	0.165	0.166	1.9
6.25	0.318	0.309	0.312	0.307	0.312	1.5
12.5	0.613	0.604	0.600	0.594	0.603	1.3
25	1.144	1.118	1.125	1.124	1.128	1.0
50	2.068	2.037	2.035	2.048	2.047	0.7

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.5	6.1-11.4
B	6.0	4.4-8.2
C	3.2	2.8-5.3