



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

February 17, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	22FESFOA129	FEB.16 2023

The product kit, batch 22FESFOA129, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 17, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.007	0.007	0.008	0.007	
0.78	0.045	0.044	0.045	0.047	0.045	2.8
1.56	0.082	0.077	0.078	0.078	0.079	2.8
3.13	0.157	0.152	0.148	0.150	0.152	2.5
6.25	0.299	0.289	0.282	0.287	0.289	2.5
12.5	0.572	0.553	0.552	0.560	0.559	1.6
25	1.109	1.086	1.069	1.068	1.083	1.8
50	2.029	1.951	1.969	1.980	1.982	1.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.3	6.1-11.4
B	6.5	4.4-8.2
C	3.7	2.8-5.3