



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

September 28, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	21SESFOA120	SEP.27 2022

The product kit, batch 21SESFOA120, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 28, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.009	0.008	0.008	0.007	0.008	
0.78	0.049	0.049	0.046	0.048	0.048	2.9
1.56	0.086	0.084	0.082	0.084	0.084	1.9
3.13	0.166	0.161	0.158	0.162	0.162	2.0
6.25	0.321	0.311	0.304	0.310	0.312	2.3
12.5	0.614	0.606	0.595	0.592	0.602	1.7
25	1.153	1.151	1.117	1.134	1.139	1.5
50	2.106	2.119	2.021	2.069	2.079	2.1

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.1	6.1-11.4
B	5.6	4.4-8.2
C	3.9	2.8-5.3