



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 20, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	21MYSFOA114	MAY.19 2022

The product kit, batch 21MYSFOA114, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 20, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.009	0.008	0.005	0.007	0.007	
0.78	0.044	0.043	0.044	0.045	0.044	1.9
1.56	0.077	0.080	0.082	0.078	0.079	2.8
3.13	0.147	0.151	0.151	0.154	0.151	1.9
6.25	0.283	0.293	0.304	0.293	0.293	2.9
12.5	0.554	0.566	0.571	0.582	0.568	2.0
25	1.042	1.040	1.044	1.077	1.051	1.7
50	1.912	1.924	2.035	1.911	1.946	3.1

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.7	6.1-11.4
B	5.5	4.4-8.2
C	4.1	2.8-5.3