

Certificate of Analysis

Product EC Blue 100 color comparator II (Code 06514)

Lot No 014404

Expiry Date 2025 . 1

	Specification	Results
Appearance	Clear greenish blue solution. No particular matters. Under 365nm UV lamp, fluorescence is equivalent to control.	Pass
Absorbance	0.07~0.12 (at 610 nm)	0 . 0 9
pH	7.2 ~ 7.6	7 . 4

Judge: Acceptable quality.

Certificate issue date :2024 . 4 . 16

Signed (QC Manager)



Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611