



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

January 14, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	21JASFOA105	JAN.13 2022

The product kit, batch 21JASFOA105, passed the quality control.

(1) Absorbance of the standard curve

(Assay date :January 14, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.017	0.018	0.016	0.017	0.017	
0.78	0.049	0.049	0.047	0.049	0.049	2.1
1.56	0.086	0.084	0.083	0.082	0.084	2.0
3.13	0.158	0.152	0.151	0.148	0.152	2.8
6.25	0.300	0.295	0.287	0.280	0.291	3.0
12.5	0.573	0.555	0.552	0.533	0.553	3.0
25	1.075	1.061	1.023	1.008	1.042	3.0
50	1.934	1.970	1.905	1.838	1.912	2.9

(2)Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.6	6.1-11.4
B	5.9	4.4-8.2
C	3.8	2.8-5.3