



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

December 23, 2020

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	20DESFOA104	DEC.22 2021

The product kit, batch 20DESFOA104, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date :December 23, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.015	0.015	0.014	0.014	0.015	
0.78	0.050	0.048	0.052	0.047	0.050	4.5
1.56	0.085	0.084	0.085	0.082	0.085	1.7
3.13	0.154	0.152	0.153	0.152	0.154	0.6
6.25	0.298	0.292	0.287	0.287	0.298	1.8
12.5	0.574	0.559	0.557	0.540	0.574	2.5
25	1.089	1.062	1.040	1.032	1.089	2.4
50	1.995	1.966	1.928	1.938	1.995	1.5

### (2)Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.1	6.1-11.4
B	6.1	4.4-8.2
C	4.1	2.8-5.3