



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

May 15, 2020

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	20MYSFOA095	MAY.12 2021

The product kit, batch 20MYSFOA095, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : May 15, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.010	0.009	0.010	0.010	
0.78	0.047	0.047	0.047	0.046	0.047	1.1
1.56	0.085	0.085	0.085	0.085	0.085	0.0
3.12	0.161	0.157	0.160	0.158	0.159	1.1
6.25	0.313	0.309	0.309	0.309	0.310	0.6
12.5	0.600	0.586	0.596	0.592	0.594	1.0
25	1.141	1.122	1.133	1.133	1.132	0.7
50	2.112	2.049	2.112	2.078	2.088	1.5

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.7	6.1-11.4
B	6.2	4.4-8.2
C	4.5	2.8-5.3