



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL: +81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

April 10, 2020

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	20APSFOA093	APR.09 2021

The product kit, batch 20APSFOA093, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : April 10, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.012	0.012	0.012	0.012	
0.78	0.050	0.048	0.049	0.047	0.049	2.7
1.56	0.091	0.086	0.088	0.086	0.088	2.7
3.12	0.173	0.165	0.162	0.162	0.166	3.1
6.25	0.328	0.308	0.316	0.311	0.316	2.8
12.5	0.633	0.598	0.616	0.597	0.611	2.8
25	1.209	1.150	1.165	1.123	1.162	3.1
50	2.169	2.102	2.089	2.119	2.120	1.7

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.0	6.1-11.4
B	6.0	4.4-8.2
C	4.2	2.8-5.3