



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

February 13, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	20FESFOA091	FEB.12 2021

The product kit, batch 20FESFOA091, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 13, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.012	0.013	0.012	0.012	4.1
0.78	0.048	0.047	0.046	0.046	0.047	2.0
1.56	0.085	0.080	0.081	0.080	0.082	2.9
3.12	0.157	0.152	0.151	0.147	0.152	2.7
6.25	0.303	0.297	0.290	0.284	0.294	2.8
12.5	0.606	0.571	0.574	0.559	0.578	3.5
25	1.127	1.091	1.093	1.067	1.095	2.3
50	2.054	1.997	1.937	1.942	1.983	2.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.4	6.1-11.4
B	6.5	4.4-8.2
C	4.0	2.8-5.3