

November 6, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	19NOSFOA088	NOV.05 2020

The product kit, batch 19NOSFOA088, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : November 6, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.008	0.009	0.007	0.009	15.2
0.78	0.049	0.050	0.048	0.051	0.050	2.6
1.56	0.089	0.094	0.091	0.097	0.093	3.8
3.12	0.173	0.176	0.176	0.181	0.177	1.9
6.25	0.332	0.340	0.328	0.345	0.336	2.3
12.5	0.636	0.658	0.645	0.649	0.647	1.4
25	1.232	1.235	1.237	1.263	1.242	1.2
50	2.176	2.210	2.197	2.207	2.198	0.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.7	6.1-11.4
B	5.8	4.4-8.2
C	3.8	2.8-5.3