



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

June 06, 2019

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	19JUSFOA080	JUN.04 2020

The product kit, batch 19JUSFOA080, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : June 06, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.025	0.019	0.021	0.018	0.021	14.9
0.78	0.068	0.057	0.063	0.060	0.062	7.6
1.56	0.100	0.093	0.096	0.094	0.096	3.2
3.12	0.177	0.166	0.171	0.170	0.171	2.7
6.25	0.319	0.311	0.314	0.309	0.313	1.4
12.5	0.603	0.593	0.599	0.591	0.597	0.9
25	1.133	1.131	1.125	1.127	1.129	0.3
50	2.065	2.063	2.054	2.044	2.057	0.5

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.4	6.1-11.4
B	6.8	4.4-8.2
C	4.4	2.8-5.3