



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

April 03, 2019

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Egg(Ovalbumin) ELISA Kit II	19APSFOA079	APR.02 2020

The product kit, batch 19APSFOA079, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : April 03, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.014	0.015	0.012	0.013	0.014	9.6
0.78	0.061	0.054	0.058	0.054	0.057	6.0
1.56	0.097	0.092	0.091	0.086	0.092	4.9
3.12	0.172	0.160	0.171	0.152	0.164	5.8
6.25	0.324	0.307	0.313	0.301	0.311	3.2
12.5	0.608	0.584	0.574	0.565	0.583	3.2
25	1.154	1.103	1.071	1.061	1.097	3.8
50	2.098	1.993	1.969	1.955	2.004	3.2

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.3	6.1-11.4
B	6.2	4.4-8.2
C	4.1	2.8-5.3