



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

July 25, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

### Product:

Product name	Batch number	Expiration date
Casein ELISA Kit II	22JLSFCS119	JUL.22 2023

The product kit, batch 22JLSFCS119, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date: July 26, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.011	0.009	0.011	0.010	
0.78	0.053	0.046	0.050	0.050	0.050	5.8
1.56	0.090	0.089	0.087	0.087	0.088	1.7
3.13	0.164	0.158	0.162	0.159	0.161	1.7
6.25	0.312	0.298	0.305	0.301	0.304	2.0
12.5	0.587	0.572	0.572	0.566	0.574	1.6
25	1.100	1.097	1.094	1.072	1.091	1.2
50	2.007	1.975	1.973	1.983	1.985	0.8

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.7	8.5-15.7
B	6.8	5.4-10.0
C	6.3	5.4-10.0