



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

January 12, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Casein ELISA Kit II	22JASFCS110	JAN.11 2023

The product kit, batch 22JASFCS110, passed the quality control.

(1) Absorbance of the standard curve

(Assay date: January 12, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.040	0.038	0.036	0.038	0.038	
0.78	0.089	0.085	0.084	0.084	0.086	2.8
1.56	0.132	0.129	0.127	0.125	0.128	2.3
3.13	0.215	0.206	0.205	0.206	0.208	2.3
6.25	0.385	0.369	0.361	0.364	0.370	2.9
12.5	0.700	0.672	0.666	0.652	0.673	3.0
25	1.264	1.238	1.215	1.210	1.232	2.0
50	2.230	2.216	2.174	2.166	2.197	1.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	11.3	8.5-15.7
B	6.5	5.9-10.9
C	6.7	4.9-9.0