



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

April 16, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Casein ELISA Kit II	20APSFCS084	APR.15 2021

The product kit, batch 20APSFCS084, passed the quality control.

(1) Absorbance of the standard curve

(Assay date: April 16, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.002	0.004	0.004	0.003	0.003	
0.78	0.038	0.039	0.039	0.038	0.039	1.5
1.56	0.076	0.078	0.076	0.075	0.076	1.7
3.12	0.150	0.166	0.157	0.152	0.156	4.6
6.25	0.300	0.303	0.310	0.302	0.304	1.4
12.5	0.685	0.592	0.609	0.657	0.636	6.7
25	1.131	1.131	1.132	1.094	1.122	1.7
50	2.002	2.011	2.030	2.027	2.018	0.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	13.3	8.5-15.7
B	8.6	5.9-10.9
C	7.7	4.9-9.0