



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

June 05, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Casein ELISA Kit II	19JUSFCS074	JUN.04 2020

The product kit, batch 19JUSFCS074, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 05, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.008	0.007	0.008	0.008	7.7
0.78	0.042	0.045	0.044	0.046	0.044	3.9
1.56	0.078	0.077	0.076	0.079	0.078	1.7
3.12	0.151	0.148	0.141	0.148	0.147	2.9
6.25	0.286	0.275	0.278	0.266	0.276	3.0
12.5	0.551	0.545	0.543	0.530	0.542	1.6
25	1.061	0.992	0.992	0.979	1.006	3.7
50	1.871	1.921	1.828	1.845	1.866	2.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.1	8.5-15.7
B	6.9	5.9-10.9
C	6.4	4.9-9.0