

May 19, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Buckwheat ELISA Kit II	22MYSFBW068	MAY.18 2023

The product kit, batch 22MYSFBW068, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 19, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.009	0.008	0.007	0.008	0.008	
0.78	0.049	0.048	0.046	0.047	0.048	2.7
1.56	0.090	0.086	0.085	0.085	0.087	2.8
3.13	0.170	0.167	0.163	0.159	0.165	2.9
6.25	0.325	0.317	0.310	0.309	0.315	2.4
12.5	0.618	0.614	0.591	0.598	0.605	2.1
25	1.176	1.147	1.143	1.128	1.149	1.7
50	2.139	2.094	2.102	2.064	2.100	1.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	17.4	11.3-21.0
B	16.1	11.1-20.6
C	8.1	5.7-10.6