



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,

YOKOHAMA, 2360003, JAPAN

TEL:+81-45-791-7674

E-mail: info_miobs_e@morinaga.co.jp

November 26, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Buckwheat ELISA Kit II	20NOSFBW056	NOV. 24 2021

The product kit, batch 20NOSFBW056, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : November 26, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.014	0.014	0.015	0.015	0.015	
0.78	0.060	0.058	0.058	0.056	0.058	2.8
1.56	0.105	0.104	0.102	0.103	0.104	1.2
3.13	0.189	0.190	0.187	0.192	0.190	1.1
6.25	0.364	0.361	0.355	0.365	0.361	1.2
12.5	0.695	0.682	0.679	0.688	0.686	1.0
25	1.281	1.250	1.255	1.279	1.266	1.3
50	2.263	2.218	2.219	2.277	2.244	1.3

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	13.2	11.3-21.0
B	14.4	11.1-20.6
C	7.7	5.7-10.6