



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,

YOKOHAMA, 2360003, JAPAN

TEL:+81-45-791-7674

E-mail: info\_miobs\_e@morinaga.co.jp

January 8, 2020

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Buckwheat ELISA Kit II	20JASFBW049	JAN.07 2021

The product kit, batch 20JASFBW049, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : January 8, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.018	0.023	0.017	0.018	0.019	14.3
0.78	0.065	0.064	0.063	0.061	0.063	2.7
1.56	0.105	0.102	0.103	0.102	0.103	1.4
3.12	0.189	0.186	0.189	0.181	0.186	2.0
6.25	0.339	0.344	0.345	0.331	0.340	1.9
12.5	0.646	0.655	0.646	0.627	0.644	1.8
25	1.206	1.219	1.216	1.191	1.208	1.0
50	2.157	2.165	2.172	2.164	2.165	0.3

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	13.3	11.3-21.0
B	13.8	11.1-20.6
C	7.5	5.7-10.6