

# CERTIFICATE OF ANALYSIS

## CheckLite HS set

Product code :61310

Lot No.: 20230712A

Expiry date: Nov. 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Luciferin-luciferase reagent HS Luminescence:Std.(%)	98.4	100±15
Reconstitution buffer for luciferin-luciferase reagent Background ATP( $\times 10^{-13}$ M)	0.6	$\leq 4.0$
ATP releasing reagent Background ATP( $\times 10^{-13}$ M)	0.1	$\leq 4.0$
ATP eliminating reagent Adenosine phosphate deaminase activity(U/vial)	24.2	$\geq 20$
Reconstitution buffer for ATP eliminating reagent Background ATP( $\times 10^{-13}$ M)	$\leq 0.0$	$\leq 4.0$

Storage condition : at 2-8°C protected from light

  
Ryosuke Sugimoto; Manager

Date Jul. 24, 2023

KBG6031501af-Eaa  
SA2307-0239

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company