

CERTIFICATE OF ANALYSIS

CheckLite HS set

Product code :61310

Lot No.: 20231006A

Expiry date: Feb. 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Luciferin-luciferase reagent HS Luminescence:Std.(%)	98.4	100±15
Reconstitution buffer for luciferin-luciferase reagent Background ATP($\times 10^{-13}$ M)	0.6	≤ 4.0
ATP releasing reagent Background ATP($\times 10^{-13}$ M)	0.1	≤ 4.0
ATP eliminating reagent Adenosine phosphate deaminase activity(U/vial)	24.2	≥ 20
Reconstitution buffer for ATP eliminating reagent Background ATP($\times 10^{-13}$ M)	≤ 0.0	≤ 4.0

Storage condition : at 2-8°C protected from light


Ryosuke Sugimoto; Manager

Date Oct. 23, 2023

KBG6031501af-Eaa
SA2310-0247

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company