

CERTIFICATE OF ANALYSIS**CheckLite 250 Plus**

Product code :61305

Lot No.: 20231206A

Expiry date: Apr. 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Luciferin-luciferase reagent Luminescence:Std.(%)	97.7	100±15
Reconstitution buffer for luciferin-luciferase reagent Background ATP($\times 10^{-13}$ M)	0.8	≤ 4.0
ATP releasing reagent Background ATP($\times 10^{-13}$ M)	0.1	≤ 4.0

Storage condition : at 2-8°C, protected from light

This is to certify that above information is a true accurate copy of data obtained for this kit.


Ryosuke Sugimoto; ManagerDate Dec. 20, 2023KBG6031201ae-Eaa
SA2312-0237Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company