

# CERTIFICATE OF ANALYSIS

## CheckLite 250 Plus

Product code :61305

Lot No.: 20231206B

Expiry date: Apr. 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Luciferin-luciferase reagent Luminescence:Std.(%)	100.0	100±15
Reconstitution buffer for luciferin-luciferase reagent Background ATP( $\times 10^{-13}$ M)	0.8	$\leq 4.0$
ATP releasing reagent Background ATP( $\times 10^{-13}$ M)	0.1	$\leq 4.0$

Storage condition : at 2-8°C, protected from light

This is to certify that above information is a true accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Dec. 20, 2023

KBG6031201ae-Eaa  
SA2312-0238

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company