



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac Pen-AQUA

Product code :60336

Lot No.: 20230831Q

Expiry date: Nov. 30, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Oct. 16, 2023

KBG6033600ab-Eaa
SA2310-0150

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company