



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen-AQUA**

Product code :60336

Lot No.: 20240115Q

Expiry date: Apr. 14, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Feb. 19, 2024

KBG6033600ab-Eaa  
SA2402-0187

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company