



CERTIFICATE OF ANALYSIS

LuciPac Pen-AQUA

Product code :60336


Lot No.: 20240426Q

Expiry date: Jul. 25, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$

(When applied 100 μ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date May. 20, 2024

KBG6033600ab-Eaa
SA2405-0200

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company