



**CERTIFICATE OF ANALYSIS**

**LuciPac Pen**


Product code :60331

Lot No.: 20230703U

Expiry date: Oct. 02, 2024

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$
(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; Manager

Date Jul. 24, 2023

KBG6033100ab-Eaa  
SA2307-0226

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company